

ENTRY NO. 6

NAME OF MACHINE IPEN-CNEN/SP - Compact Cyclotron - CV - 28
 INSTITUTION Instituto de Pesquisas Energéticas e Nucleares - CNEN/SP
 ADDRESS C.P- 11049 - Cidade Universitária - São Paulo - Brasil - CEP- 01000
 TEL 211-6011 TELEX (011) 23592
 IN CHARGE G. Lucki REPORTED BY G. Lucki

HISTORY AND STATUS

DESIGN, date 1970 Model tests 1976
 ENG DESIGN, date 1970
 CONSTRUCTION, date 1975
 FIRST BEAM, date (or goal) 1981
 MAJOR ALTERATIONS none

COST, ACCELERATOR US\$ 1.10⁶
 COST, FACILITY, total US\$ 2 X 10⁶
 FUNDED BY Brazilian Government

ACCELERATOR STAFF, OPERATION AND DEVELOPMENT

SCIENTISTS 3 ENGINEERS 2
 TECHNICIANS 2 CRAFTS 1
 GRAD STUDENTS involved during year -
 OPERATED BY - Research staff or 5 Operators
 OPERATION 20 hr/wk. On target 3-5 hr/wk
 TIME DISTR. in house 100 % outside 0 %
 BUDGET, op & dev US\$ 100.000
 FUNDED BY Brazilian Government

RESEARCH STAFF, not included above

USERS, in house 6 outside none
 GRAD STUDENTS involved during year -
 RESEARCH BUDGET, in house -
 FUNDED BY -

MAGNET

POLE FACE, diameter (compact) 96 cm, R-extraction 42 cm
 R injection 1 cm
 GAP, min 5.1 cm, Field 14 kG }
 max 11.6 cm, Field 21 kG } at 0.23 - 10⁶
 AVERAGE FIELD at R ext 17 kG } Ampere turns
 B max / < B > 1.2
 NUMBER OF SECTORS { compact 3 } Spiral, max deg
 separated 3 }
 SECTOR ANGLE (SSC) 60 deg
 TRIMMING COILS 4

CONDUCTOR, material and type Cu
 STORED ENERGY (cryogenic) MJ
 POWER: main coils 65 max kW; current stability 50uA
 trimming coils 54 max kW; current stability 50uA
 WEIGHT: Fe 20.5 tons; coils ~2.0 tons
 COOLING system Demineralized Water
 ION ENERGY (Bending limit) E/A = q²/A² MeV/amu
 (Focusing limit) E/A = q/A MeV/amu

ACCELERATION SYSTEM

DEES, number 2 angle 90 deg
 BEAM APERTURE 2.5 cm; DC Bias 1.5-2.5 kV
 TUNED by, coarse fine
 RF 6.6 to 25.9 MHz, stable ± Better 0.01%
 Orb F 6.6 to 25.9 MHz
 HARMONICS, RF/Orb F, used fundamental
 DEE-Gnd, max kV, min gap cm
 STABILITY, (pk-pk noise)/(pk RF volt) 1%
 ENERGY GAIN, max kV/turn
 RF PHASE, stable to ± deg
 RF POWER input, max 65 kW
 FREQUENCY MODULATION, rate /s
 modulator, type
 beam pulse, width

VACUUM SYSTEM

OPERATING PRESSURE 5.10⁻⁵ Torr or mbar
 PUMPS, No, Type, Size 1 Diff. pump 12"

ION SOURCES

PIG or Thermoionic

INJECTION SYSTEM

Cold cathode - PIG or THERMIONIC

EXTRACTION SYSTEM

Electrostatic bar + Magnetic Channel

FACILITIES FOR RESEARCH

SHIELDED AREA, fixed 80 m²; movable 180 m²
 TARGET STATIONS 2 in 2 rooms
 STATIONS served at same time, max 1 (one)
 MAG SPECTROGRAPH, type
 COMPUTER model
 OTHER FACILITIES cooling system

CHARACTERISTIC BEAMS

PARTICLE	ENERGY (MeV)		CURRENT (pA)	
	Goal	Achieved	Internal	External
p	24	24	300	40
α	28	28	50	25
He ³	36	36	50	25
d	14	14	3.0	0.5

SECONDARY (part/s)

BEAM PROPERTIES

MEASURED CONDITIONS
 PULSE WIDTH RF deg pA of MeV ions
 PHASE EXC, max RF deg pA of MeV ions
 EXTRACT eff 75 % max 30 pA of 24 MeV p ions
 RESOL ΔE/E 0.5 % pA of MeV ions
 EMITTANCE
 (π mm-mrad) 30 axial pA of MeV
 30 rad

OPERATING PROGRAMS, time distribution

BASIC NUCLEAR PHYSICS - SOLID STATES PHYSICS -
 BIOMEDICAL APPLICAT - ISOTOPE PRODUCTIONS 100%

REFERENCES/NOTES

- Lucki G; et al.
- 4 th - Japan - Brasil Symposium on Science and Technology. - São Paulo Aug. 1984.

PLAN VIEW OF FACILITY, COMMENTS, ETC.

