

ENTRY NO. FM-3

NAME OF MACHINE SYNCHRO-CYCLOTRON LYON
 INSTITUTION Service Commun. du Synchrocyclotron
 ADDRESS Université Claude Bernard Lyon-1-43 Bd. du 11. novembre 1918 - 69622 Villeurbanne Cedex (France)
 TEL 78 89 81 24 TELEX IPN Lyon 380273F
 IN CHARGE Prof. B.E. LAHNECHE REPORTED BY G. HADINGER

HISTORY AND STATUS

DESIGN, date Model tests
 ENG DESIGN, date 1956
 CONSTRUCTION, date 1962
 FIRST BEAM, date (or goal) 1963
 MAJOR ALTERATIONS 1965

COST, ACCELERATOR 2,7 MF
 COST, FACILITY, total 7,5 MF
 FUNDED BY Ministre de l'Education Nationale

ACCELERATOR STAFF, OPERATION AND DEVELOPMENT

SCIENTISTS 1 ENGINEERS 1
 TECHNICIANS 2 CRAFTS
 GRAD STUDENTS involved during year
 OPERATED BY Research staff or 2 Operators
 OPERATION hr/wk. On target hr/wk
 TIME DISTR. in house %, outside %
 BUDGET, op & dev
 FUNDED BY

RESEARCH STAFF, not included above

USERS, in house outside
 GRAD STUDENTS involved during year
 RESEARCH BUDGET, in house
 FUNDED BY

MAGNET

POLE FACE, diameter (compact) 180 cm, R-extraction 75 cm
 R injection cm
 GAP, min 35 cm, Field 14.7 kG
 max 35 cm, Field 14.7 kG at 0.61 · 10⁶
 AVERAGE FIELD at R ext kG Ampere turns
 B max / < B >

NUMBER OF SECTORS {compact separated} Spiral, max deg
 SECTOR ANGLE (SSC) deg
 TRIMMING COILS

CONDUCTOR, material and type ALUMINIUM
 STORED ENERGY (cryogenic) 1 MJ
 POWER: main coils max kW: current stability
 trimming coils max kW: current stability

WEIGHT: Fe 200 tons: coils tons
 COOLING system

ION ENERGY (Bending limit) E/A = q²/A² MeV/amu
 (Focusing limit) E/A = q/A MeV/amu

ACCELERATION SYSTEM

DEES, number 2 angle 180 deg
 BEAM APERTURE 18,5 cm; DC Bias 0.2 kV
 TUNED by, coarse fine
 RF 10.4 to 11.0 MHz, stable ±
 Orb F to MHz

HARMONICS, RF/Orb F, used
 DEE-Gnd, max 22 kV, min gap cm
 STABILITY, (pk-pk noise)/(pk RF volt)
 ENERGY GAIN, max kV/turn
 RF PHASE, stable to ± deg
 RF POWER input, max 23 kW

FREQUENCY MODULATION, rate 2.000 /s
 modulator, type Rotating capacitor
 beam pulse, width 40 μsec macrocycle

VACUUM SYSTEM

OPERATING PRESSURE 8x10⁻⁶ Torr or mbar
 PUMPS, No, Type, Size 1 Diffusion pump
 50 cm diamètre

ION SOURCES

OPEN ION SOURCE

INJECTION SYSTEM**EXTRACTION SYSTEM**

Magnetic Regenerative

FACILITIES FOR RESEARCH

SHIELDED AREA, fixed 160 m²; movable m²
 TARGET STATIONS 6 in 2 rooms
 STATIONS served at same time, max
 MAG SPECTROGRAPH, type
 COMPUTER model
 OTHER FACILITIES

CHARACTERISTIC BEAMS

PARTICLE	ENERGY (MeV)		CURRENT (pμA)	
	Goal	Achieved	Internal	External
d		28	30	1.5
α		56	1.0	
SECONDARY			(part/s)	

BEAM PROPERTIES

MEASURED	CONDITIONS	
	RF deg	μA of MeV ions
PULSE WIDTH		
PHASE EXC, max		
EXTRACT eff		
RESOL ΔE/E2		
EMITTANCE		
(π mm-mrad)	40, axial	μA of MeV
	40, rad	

OPERATING PROGRAMS, time distribution

BASIC NUCLEAR PHYSICS SOLID STATES PHYSICS
 BIOMEDICAL APPLICAT. 35% ISOTOPE PRODUCTIONS 30%
 TEST OF EXPERIMENTAL DEVICES 35%

REFERENCES/NOTES

- 1)
- 2)

PLAN VIEW OF FACILITY, COMMENTS, ETC.