

ITAR BARRARE STATE

# Modeling and Performance

# **Evaluation of DCCTs in SSRF**

#### **Zhichu CHEN**

On behalf of Beam Diagnostics group of SSRF Shanghai Institute of Applied Physics

#### **Outline**



- Introduction to SSRF
- Overview of the DCCT system on the SR
- Noise Study
  - ➤ characterization
  - $\succ$  noise exploration
  - ➤ modeling
  - ➤ workaround
  - ➤ result
- Future tests
- Summary



# Introduction to SSRF

Zhichu CHEN / SSRF BI group

#### **Introduction to SSRF**



#### 3.5GeV storage ring



	DBA	Low Emittance mode	Normal mode
Energy	GeV	3.5	3.5
Circumference	m	432	432
Natural Emittance	nm*rad	3.9	11.2
Current	mA	200-300/5	200-300/5
Tune		22.22/11.32	22.22/11.32
Chromaticity		-56/-19	-56/-19
RF Voltage	MV	4.0~6.0	4.0~6.0
Energy Loss/Turn	MeV	1.448	1.448
Max beam power	kW	~600	~600

#### **150MeV LINAC**

The only third generation light source in China 2465 proposals, 235 research teams, 3780 users since formal operation in 2009 Availability 97.6%, MTBF 55.3 hours, MDT 1.55 hours in 2011



# **Overview of the DCCT**

Zhichu CHEN / SSRF BI group

#### **The Primary Requirements**





At least 2 DCCTs required:

**Display current details during ramping cycle @ Booster** (whole ramping cycle 250ms, average current < 2mA)

Precise average current & lifetime measurement @ Ring (typical current [140mA 210mA], lifetime 20~30hours)

	Design goals
Ring DCCT	10 μ A @ 1Hz
Booster DCCT	50 μ A @ 10kHz

A general solution adopted to cover two applications:

Raw sampling rate 10kHz, to capture booster current waveform

The 2Hz high resolution data were produced by averaging the 10kHz data at the PC end Both 10kHz and 2Hz data available to users

Total 2 DCCTs installed in the Ring DCCT15, primary average current monitor DCCT17, online backup

## **System Configuration**



Component	Solution	
Sensor	NPCT175 HR, 1 $\mu$ A/ $\sqrt{Hz}$	
Shielding	Fabrication locally, design borrowed from SPEAR3	
DAQ	PXI NI4070, 6.5bits DVM, 10kHz sampling rate	
Data processing	digital filtering + averaging	
Data server	10kHz and 2Hz data, EPICS IOCcore @ Linux	



## **Shielding Design**





design





Noise test has performed for: NI PXI 4070 (6.5bit) NI PXI 4071 (7.5bit) NI PXI 4070 combined with NPCT175



- 4071 do not have a significantly better performance
  - The measurement noise mainly comes from NPCT sensor under the sampling rate of 40kHz and less
  - The main source becomes the DVM electronics when the sampling rate is above 40kHz

#### **System Performance**





DCCT15 is better than DCCT17 (installation, ground loop, etc.)

Resolution of 2Hz data is close to the hardware limitation, hard to improve further

Various noise have been found at 10kHz data, lots of room for optimization

# SSRF Ring DCCT linearity test @ 2009.03.31

Nonlinearity < 0.03% @ 0 ~ 300mA



# **Noise Study**

Zhichu CHEN / SSRF BI group

## **Noise Study: Characterization**



Type of noise	frequency	amplitude	occurrence
power line noise	50Hz and its harmonics	<10 µ A for DCCT15	always
frequency specific noise	1.147kHz for DCCT15	~300 µ A	always
frequency drifting noise	few kHz	~100 µ A	always
square wave noise	/	-600 μ A	~once a week
zero drift	white	<1 µ A@1Hz	always





#### **Square Wave Noise: Characteristic**



It hurts the resolution badly, especial for lifetime measurement

Hard to be removed with traditional timedomain or frequencydomain filtering means

104 103.8 103.6 103.4 103.2 103.2 103.2 103.2 103.4 102.8 102.6 1 2 3 4 5 6 7 8

appear **randomly** without any warning

shaped as a square waveform with indefinite periods and fixed amplitudes

#### **Narrowband Noise : Characteristic**





- power line noise (50Hz) and its harmonics
- frequency specific noise (1.147kHz)
- frequency drifting noise(4.4kHz—2.2kHz during this period)
- The total intensity of the narrowband noise is approximately proportional to the beam current



## **Noise Study: Influence**

	-	-	
-	2.0		
	20		
TI V	Sec. 2		

Type of noise	influence
power line noise	the resolution of the 10kHz data
frequency specific noise	
frequency drifting noise	
square wave noise	resolution of both the 10kHz data and the 2Hz data; system measurement error; lifetime measurement error
zero drift	resolution of both the 10kHz data and the 2Hz data

## **Noise Study: Exploration**

1000	-	
		-

Type of noise	exploration
power line noise	feed reverse current to cancel out the DC component of the
frequency specific noise	beam current, to see whether it comes from: sensor, nonlinear
frequency drifting noise	ADC, etc.
square wave noise	try to find the correlation between the noise and other machine parameters: RF voltages, beam current, etc.
zero drift	/

## **Narrowband Noise Exploration**

-		-
	ala to	- Andrewson
	2000	

Type of noise	exploration
power line noise	feed reverse current to cancel out the DC component of the
frequency specific noise	beam current, to see whether it comes from: sensor, nonlinear ADC, etc.
frequency drifting noise	



- Most of the NB noise is excited by the beam
- The spectrum scarcely changed after the reverse current was fed. Hence, the noise is DC unrelated.



#### Noise spectrum without beam, with test signal and with beam



Noise spectrum with & without inverse current

## **Square Wave Noise: Exploration**

Type of noise	exploration
square wave noise	try to find the correlation between the noise and other machine
	parameters: RF voltages, beam current, etc.



No other relations have been found between the noise and machine parameters except that this type of noise does not appear when the beam current is below 197mA, and has an interesting distribution around the high current area





## **Noise Study: Workaround**



Type of noise	workaround
power line noise	narrowband, multiple frequencies, unstable frequencies:
frequency specific noise	adaptive notch filter
frequency drifting noise	
square wave noise	unpredictable occurrences, wideband, hard to remove in time- domain as well as in frequency-domain: Kalman filter
zero drift	installation optimization



#### Workaround: Adaptive Notch Filter



current waveform before and after filtering



histogram before and after filtering



spectrum before and after filtering

original signal FFT spectrum frequency detector time-domain filter threshold noise detector found signal qualified output signal





std(filtered)=0.053mA std(kalman)=0.006mA

#### Zhichu CHEN / SSRF BI group

#### **Noise Study: Result**



Type of noise	result
power line noise	resolution reduces from (DCCT15/DCCT17 @10kHz)
frequency specific noise	80 μ A~116 μ A/184 μ A~269 μ A
frequency drifting noise	to 12.4 μ A/30 μ A
square wave noise	resolution reduces from 53 $\mu$ A to 6 $\mu$ A for 2Hz data
zero drift	/



Resolution before and after adaptive notch filtering @ 5kHz BW



Resolution before and after Kalman filtering @ 1Hz BW



•No significant improvement for 2Hz data

•Much better than the specification for 10kHz sampling rate after the filtering

•Improved 10kHz data could be useful for machine study (such as energy calibration)



# **Future Tests**

Zhichu CHEN / SSRF BI group

- Make more correlation analysis to locate the source of the square wave noise
- Optimize the installation to minimize the zero drift
  - The 10kHz data (on the storage ring) is filtered in order to increase the 2Hz data performance, which turns out it's limited by the zero drift
- Use BPM sum signals as a replacement for some special situations
  - Some algorithms have already been tested to pick up the "good" BPMs as talked in another article of these proceedings.
- And more . . .



# Summary

Zhichu CHEN / SSRF BI group

#### Conclusions

- CONTRACTOR
- The combination of NPCT175+NI 4070 DVM with sampling rate of 10kHz was picked up to meets the basic needs in SSRF
- Various kinds of noise exist and affect the system despite the shielding, including NB noise, random square wave from nowhere and zero drift problem
- After using the adaptive notch filter and the Kalman filter, the performance of the DCCT has a significant improvement
  - ➤ Adaptive notch filter (DCCT15/DCCT17 @10kHz):
    - 80  $\mu$  A~116  $\mu$  A/184  $\mu$  A~269  $\mu$  A to 12.4  $\mu$  A/30  $\mu$  A
  - ➢ Kalman filter (@2Hz):
    - 53 µ A to 6 µ A
- Anyway, the methods are intended only as a stopgap. We will keep looking for where the noise comes from and try to get rid of it from the first place