



Systematic Errors

At start of calibration procedure accuracy of measurements was estimated by measure only cables without I-Q modulator. That is defining threshold of iterations cycles and bottom line of the systematic error. For power it was found as  $\delta A \leq 0.005 \text{ dB}$ , for phase  $\delta \varphi \leq 0.07 \text{ deg}$ . After that I-Q modulator was connected and whole measurement cycle of 360 points at constant power (20dB) and phase (100deg) settings was repeated. This is define accuracy of I-Q modulator itself and was found as  $\Delta A \leq 0.03 \text{ dB}$ , for phase  $\Delta \varphi \leq 0.2 \text{ deg}$ .

After calibration was done, compensation tables for power and phase were built and applied for operation. Accuracy of new control circuit measured in whole operational range. It was found that most error value comes at switching power level while temperature is stabilizing in the system. This transition time takes about 5 minutes and gives error in power less than  $0.1 \text{ dB}$  and in phase less than  $0.2 \text{ deg}$ . During calibration this error was neglected by iteration cycles and overlapped by environmental temperature changes (day-night).

In the Table 1 is shown power and phase errors during scan of the most usable operational region of 0-10 dB power and 0-360 degree phase. Errors represented in rms

and max values. Since I and Q are coupled in digital 12-bit range, during scan depending on power-phase settings it could make jumps switching values from 4095 to 0 (and vice versa) that will covered by max error.

Table 1: Comparison of old and new control accuracy

	Used control	New control
<b>LINAC</b>		
Power error over all scan	rms =0.15 dB max=0.43 dB	rms=0.02 dB max=0.07 dB
Phase error over all scan	rms = 0.7 <sup>0</sup> max = 2.7 <sup>0</sup>	rms = 0.13 <sup>0</sup> max =0.57 <sup>0</sup>
<b>GUN</b>		
Power error over all scan	rms =0.14 dB max=0.48 dB	rms = 0.03 dB max= 0.10 dB
Phase error over all scan	rms = 0.85 <sup>0</sup> max = 2.73 <sup>0</sup>	rms = 0.31 <sup>0</sup> max =1.22 <sup>0</sup>




Figure 3: Old (red) and new (blue) control errors. Top pictures shows power and phase differences versus power. Bottom pictures shows difference versus phase. Scan ranges: power 0-10 dB, 1dB step; phase 0-360° with 1° step.

## SLOW FEEDBACK

When new I-Q modulator control was commissioned and put in operation, gun and linac RF control was modified for slow feedback feature. Phase meter was installed to measure phase difference between Master Oscillator input and klystrons output in both gun and linac RF circuits. It measured phase difference then use it to readjust phase set points to hold phase output close to constant. Measure - adjust cycle is going with frequency about 0.3 - 0.5 Hz and compensate all temperature dependent drifts in the system. With success of this work it become possible to carry out experiments with fine tuning beam parameters [2].

## CONCLUSION

New I-Q modulator software control was developed and commissioned at BNL ATF with operational rms errors less than 0.03dB of power and 0.3° of phase.

Slow feedback for gun and linac RF were built to compensate slow drifts that mostly have an environment temperature changes.

## REFERENCES

- [1] [www.herley.com](http://www.herley.com)
- [2] V. Yakimenko, M. Fedurin, V. Litvinenko, A. Fedotov, D. Kayran, "CSR shielding experiment", PAC11, WEP107.