

MOPSA51

THE EXTRACTION SYSTEM OF DC140 CYCLOTRON

V.Lisov[†], N. Kazarinov, A. Zabanov, I. Ivanenko, A. Protasov, K. Gikal, G. Ivanov, N. Osipov, I. Kalagin, S. Mitrofanov, V. Semin, G. Gulbekyan JINR, Dubna, Moscow region, 141980, Russia † lisov@jinr.ru







INTRODUCTION

The main activities of Flerov Laboratory of Nuclear Reactions, following its name - are related to fundamental science, but, in parallel, plenty of efforts are paid for practical applications. For the moment continues the works under creating irradiation facility based on the cyclotron DC140 which will be dedicated machine for applied researches in FLNR. The beam transport system will have three experimental beam lines for testing of electronic components (avionics and space electronics) for radiation hardness, for ion-implantation nanotechnology and for radiation materials science.

The DC140 is a sector cyclotron is intended for acceleration of heavy ions with variation of the magnetic field level at range Bo=1.415÷1.546 T [1]. The main parameters of DC140 cyclotron are given in Table 1. The DC140 will be a reconstruction of the DC72 cyclotron [2, 3]. In DC72 beam was extracted by stripping method. In DC140 the extraction will be carried out using an electrostatic deflector.



Figure 3: The horizontal and vertical beam emittances at the deflector entrance (A/Z=5.5, Bo=1.55 T)



FOCUSING MAGNETIC CHANNEL

The focusing magnetic channel (MC) is intended for focusing in the horizontal direction of the heavy ion beam extracted from the cyclotron (see Fig. 7). The MC consists of 4 parts with different cross-sections. Azimuth extent of each section 8.3° (see Fig. 8).

The main parameters of the magnetic channel are giv-en in Table 4.

 Table 4: Parameters of Magnetic Channel

Parameter	Value
Azimuthal position [deg]	36÷64
Horizontal aperture [mm]	32
Vertical aperture [mm]	20
Displace of the edges [mm]	± 15
Focusing gradient of the magnetic field [T/m]	13.5

Table 1: DC140 Cyclotron Main Parameters

Parameter	Va	lue
Pole (extraction) radius [m]	1.3(1.18)
Number of sectors	4	4
RF frequency [MHz]	8.6	532
Harmonic number	2	3
Energy [MeV/u]	4.8	2.124
A/Z range	5.0÷5.5	7.57÷8.25
RF voltage [kV]	6	0
Number of Dees	2	
Ion extraction method	electrostatic deflector	
Deflector voltage [kV]	73.5	

For beams extraction from the cyclotron is used the electrostatic deflector. The extraction system of the DC140 cyclotron consist a next elements (see Fig. 1):

. Electrostatic deflector (ESD);

- 2. Focusing magnetic channel (MC);
- 3. Permanent Magnet Quadrupole lens (PMQ).

Figure 5: The horizontal and vertical beam emittances at the PMQ entrance (A/Z=5.5, Bo=1.55 T)

ELECTROSTATIC DEFLECTOR

- The electrostatic deflector (ESD) consists of two par-allel plates (septum plate and potential plate), between which an electric field is created and deflecting the beam from the cyclotron chamber.
- The azimuthal position of electrostatic deflector is 40° (70° -



PERMANENT MAGNET QUADRUPOLE LENS

The cross-section of the permanent magnet quadrupole lens consisting of 26 permanent magnets is shown in Fig. 9. The main parameters of the permanent magnet quadrupole lens are given in Table 5.



NUMERICAL SIMULATION OF THE BEAM EXTRACTION

For numerical simulation the test ion in accordance with working diagram are used [see Fig. 2, 3, 4, 5]. The parameters of this ion are given in Table 2.



110°). Due to the low power of the accelerated ion beams, it was decided that the potential electrode will not have an active [5]. cooling system.

The main parameters of the electrostatic deflector are given in Table 3.

Table 3: Parameters of the Electrostatic Deflector

Parameter	Value
Azimuthal position [deg]	70÷110
Max. Voltage [kV]	73.5
Length of the "potencial" plate [mm]	780
Length of the "septum" plate [mm]	775
Thickness of the "septum" plate [mm]	0.3÷1.0
Material of the "potencial" plate [mm]	Al
Material of the "septum" plate [mm]	W
Gap between plates [mm]	9
Displace of the deflector edges [mm]	± 10

During the design, three modifications of the deflector were considered: with constant gap g = 9 mm, with a linearly Simulation and design of the PMQ is presented in the work

Table 5: Parameters of PMQ

Parameter	Value
Horizontal aperture [mm]	64
Vertical aperture [mm]	25
Effective length [cm]	29.9
Focusing gradient of the magnetic field [T/m]	8.1



Figure 9: Cross section of PMQ consisting of 26 permanent magnets

Figure 2: Working diagram of DC140 cyclotron

Table 2: Parameters of the Test Ions

Ion	²⁰⁹ Bi ³⁸⁺
A/Z	5.5
Bo [T]	1.5458
B(R) [T]	1.5540
E(R)	4.80
Fion [MHz]	4.316
Harmonic	2
Rmag [cm]	3
Uinj [kV]	18.86

increasing gap (g1/gn = 6/9) and with a linearly decreasing electrostatic field (g1/gn = 6/9).

Modeling of the electrostatic deflector in detail in [4].

The electrostatic deflector with constant gap between electrodes provides the lowest beam losses ($\eta = 21.25$ %) at the septum compared to the other modifications (see Fig. 6).



The working area takes into account the horizontal spread of orbits ± 4 mm.

REFERENCES

[1] S. V. Mitrofanov et al., "FLNR JINR Accelerator Complex for Applied Physics Researches: State-of-Art and Future", In Proc. of 22nd Conf. on Cycl. and their Appl., Cape Town, South Africa, Sep. 2019, p. 358-360, doi:10.18429/JACoW-CYCLOTRONS2019-FRB02

[2] B. N. Gikal et al., "Dubna Cyclotrons – Status and Plans", In Proc. of 17th Conf. on Cycl. and their Appl., Tokyo, Japan, Oct. 2004, paper 20A1, p. 100-104,

http://www.jacow.org

[3] G. Gulbekyan, I. Ivanenko, J. Franko, J. Keniz, "DC-72 Cyclotron Magnetic Field Formation", In Proc. of 19th Russian Part. Acc. Conf (RuPAC'04), Dubna, Russia, Oct. 2004, paper WENO12, p.147-49, http://www.jacow.org

[4] A. Zabanov et al., "Simulation of the electrostatic deflector of DC140 cyclotron", presented at RuPAC'21, Alushta, Crimea, Sep.-Oct. 2021, paper MOPSA48, this conference.

[5] I. Rodin "Simulation and Design of the Permanent Magnet Multipole for DC140", presented at RuPAC'21, Alushta, Crimea, Sep.-Oct. 2021, paper FRA03, this conference.